

## EU-DECLARATION OF CONFORMITY

Manufacturer Wero Swiss Med Kft.  
Ipartelep utca 6  
4220 Hajdúböszörmény  
Hungary

Single Registration Number: HU-MF-000006727

### We declare under our sole responsibility that

the medical device **Urgoacryl**  
with Basic-UDI-DI (productcode) **with REF 60643-45**  
**76305236A425EW**  
Risk Class class 1, unsterile  
According to Annex VIII, Rule 1


**meets all the provisions of the Regulation (EU) 2017 /745 which apply to it.**

Applied harmonised standards, national standards or other normative documents EN ISO 10993-1:2020  
EN ISO 15223-1:2017  
EN ISO 14971:2019

Conformity assessment procedure: Regulation (EU) 2017/745  
according to Annex IV

Place, date Hajdúböszörmény, 22.03.2022

Name and function

  
Tibor Balla, Head of Quality Management



Wero Swiss  
Med Kft.  
4220 Hajdúböszörmény, Ipartelep utca 6.  
Regisztrációs szám: 09-09-029496  
Adószám: 2630101-09